



Ohio Administrative Code

Rule 3701:1-58-93 Records of dosimetry equipment used with remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units.

Effective: August 15, 2005

(A) A licensee shall retain a record of the calibration, intercomparison, and comparisons of its dosimetry equipment done in accordance with rule 3701:1-58-60 of the Administrative Code for the duration of the license.

(B) For each calibration, intercomparison, or comparison, the record must include:

(1) The date;

(2) The manufacturer's name, model numbers and serial numbers of the instruments that were calibrated, intercompared, or compared as required by paragraphs (A) and (B) of rule 3701:1-58-60 of the Administrative Code;

(3) The correction factor that was determined from the calibration or comparison or the apparent correction factor that was determined from an intercomparison; and

(4) The names of the individuals who performed the calibration, intercomparison, or comparison.
