



## Ohio Administrative Code

### Rule 3701:1-58-94 Records of full calibrations for teletherapy units, remote afterloader units, and gamma stereotactic radiosurgery units.

Effective: October 4, 2010

---

(A) A licensee shall maintain a record of the teletherapy unit, remote afterloader unit, and gamma stereotactic radiosurgery unit full calibrations required by rules 3701:1-58-61 to 3701:1-58-63 of the Administrative Code for three years.

(B) The record must include:

(1) The date of the calibration;

(2) The manufacturer's name, model number, and serial number of the teletherapy, remote afterloader, and gamma stereotactic radiosurgery unit(s), the source(s), and the instruments used to calibrate the unit(s);

(3) The results and an assessment of the full calibrations;

(4) The results of the autoradiograph required for low dose-rate remote afterloader units; and

(5) The signature of the authorized medical physicist who performed the full calibration.

---