



Ohio Administrative Code

Rule 3701:1-58-95 Records of periodic spot-checks for teletherapy units.

Effective: August 15, 2005

(A) A licensee shall retain a record of each periodic spot-check for teletherapy units required by rule 3701:1-58-64 of the Administrative Code for three years.

(B) The record must include:

(1) The date of the spot-check;

(2) The manufacturer's name, model number, and serial number of the teletherapy unit, source and instrument used to measure the output of the teletherapy unit;

(3) An assessment of timer linearity and constancy;

(4) The calculated on-off error;

(5) A determination of the coincidence of the radiation field and the field indicated by the light beam localizing device;

(6) The determined accuracy of each distance measuring and localization device;

(7) The difference between the anticipated output and the measured output;

(8) Notations indicating the operability of each entrance door electrical interlock, each electrical or mechanical stop, each source exposure indicator light, and the viewing and intercom system and doors; and

(9) The name of the individual who performed the periodic spot-check and the signature of the authorized medical physicist who reviewed the record of the spot-check.



(C) A licensee shall retain a copy of the procedures required by paragraph (B) of rule 3701:1-58-64 of the Administrative Code until the licensee no longer possesses the teletherapy unit.
