



## Ohio Administrative Code

### Rule 3701-84-64 Equipment and supplies - pediatric intensive care service.

Effective: August 1, 2017

---

(A) Appropriate drugs for resuscitation and pediatric advanced life support shall be present and immediately available for use in the treatment of any patient in the PICU.

(B) The following life-saving, therapeutic and monitoring equipment shall be present or immediately available in the PICU:

(1) Portable equipment including:

(a) An emergency ("code" or "crash") cart;

(b) A procedure lamp;

(c) Pediatric sized blood pressure cuffs for systemic arterial pressure determination;

(d) A doppler ultrasound;

(e) An electrocardiograph;

(f) A defibrillator or cardioverter with pediatric paddles;

(g) Thermometers with a range sufficient to identify extremes of hypothermia and hyperthermia;

(h) Automated blood pressure apparatus;

(i) Transthoracic pacer with pediatric pads;

(j) Devices for accurately measuring body weight;



- (k) Cribs and beds with head pressure apparatus;
- (l) Infant warmers;
- (m) Heating and cooling blankets;
- (n) Bilirubin lights;
- (o) Temporary pacemakers;
- (p) A blood warming apparatus;
- (q) A transport monitor;
- (r) Infusion pumps with microinfusion capability;
- (s) Oxygen tanks for transport and backup;
- (t) Suction machines for transport and backup state="unchanged"/>;
- (u) Volumetric infusion pumps;
- (v) Air-oxygen blenders;
- (w) An air compressor;
- (x) Gas humidifiers;
- (y) Bag-valve mask resuscitators;
- (z) An otoscope and ophthalmoscope;
- (aa) Isolation carts; and



(bb) A portable electro-encephalogram available in the hospital for recordings.

(2) Respiratory equipment appropriate to meet the needs of all patients, including:

(a) Mechanical ventilators suitable for pediatric patients of all sizes;

(b) Pulse oximeters and transcutaneous oxygen monitors;

(c) Carbon dioxide ( $\text{CO}_2$ ) monitors;

(d) Inhalation therapy equipment;

(e) Chest physiotherapy and suctioning;

(f) Spirometers; and

(g) Continuous oxygen analyzers with alarms.

(3) Small equipment including:

(a) Tracheal intubation equipment in adequate numbers and type to intubate patients of all ages;

(b) Endotracheal tubes of all pediatric sizes;

(c) Oral/nasal airways;

(d) Flexible bronchoscope;

(e) Suction catheters;

(f) Vascular access equipment; and



(g) Surgical trays for the following::

(i) Vascular cut-downs;

(ii) Open chest procedures;

(iii) Cricothyroidectomy; and

(iv) Tracheostomy;

(h) Intraosseous (IO) needles.

(4) Bedside monitors that shall be capable of providing continuous, sufficient and appropriate monitoring that have visible and audible alarms and are capable of producing a permanent hard copy of the rhythm strip.

(C) Each provider of a PICU service shall have the capability to continuously monitor a patient's:

(1) Electrocardiogram and heart rate;

(2) Respiration;

(3) Temperature;

(4) Systemic arterial pressure;

(5) Oxygen;

(6) Carbon dioxide;

(7) Central venous pressure;

(8) Pulmonary arterial pressure;



(9) Intracranial pressure; and

(10) Four pressures simultaneously.