



## Ohio Administrative Code

### Rule 3701:1-58-85 Records of molybdenum-99, strontium-82, and strontium-85 concentrations.

Effective: October 4, 2010

---

A licensee shall maintain a record of the molybdenum-99 concentration or strontium-82 and strontium-85 concentration tests required by paragraph (B) of rule 3701:1-58-35 of the Administrative Code for three years. The record must include:

- (A) For each measured elution of technetium-99m, the ratio of the measures expressed as kilobecquerel (microcuries) of molybdenum-99 per megabecquerel (millicurie) of technetium-99m, the time and date of the measurement, and the name of the individual who made the measurement; or
  - (B) For each measured elution of rubidium-82, the ratio of the measures expressed as kilobecquerel (microcuries) of strontium-82 per megabecquerel (millicurie) of rubidium-82, kilobecquerel (microcuries) of strontium-85 per megabecquerel (millicurie) of rubidium-82, the time and date of the measurement, and the name of the individual who made the measurement.
-